

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 9492(1-3)			Exp:2023/12/31			
AccuBind® ELISA	94921		94922		94923	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Tumor Markers						
Ferritin in ng/ml	28.49	19.09-37.89	328.45	220.06-436.83	>800	
CEA in ng/ml	3.66	2.46-4.87	30.15	20.20-40.10	113.49	76.04-150.95
AFP in ng/ml	8.57	5.75-11.40	101.9	68.27-135.52	331.63	222.19-441.07
PSA in ng/ml	0.27	0.18-0.37	9.65	6.47-12.84	45.67	30.60-60.75
fPSA in ng/ml	<0.5		4.06	2.72-5.41	>10	
CA-125 in U/ml	23.14	15.50-30.77	64.83	43.43-86.22	201.04	134.69-267.38
CA-19-9 in U/ml	4.68	3.13-6.22	20.68	13.86-27.51	79.69	53.39-105.99
CEA Next Generation in ng/ml	<5		25.23	16.91-33.56	86.6	58.02-115.18
PRL in ng/ml	13.93	9.33-18.53	53.3	35.71-70.89	>100	
HCG in mIU/ml	4.89	3.28-6.50	47.42	31.77-63.07	212.09	142.10->250
TG in ng/ml	4.56	3.06-6.07	46.43	31.11-61.76	145.17	97.27-193.08
CA 15-3 in U/ml	18.91	12.30-25.53	85.65	57.38-113.91	173.57	100.28-246.85
HCG-XR in mIU/ml	5.06	3.39-6.72	53.31	35.72-70.90	233.3	156.31-310.30

DOCUMENT HISTORY			
PREPARED BY: <u>MSW</u>	DEPT: QC	VERIFIED BY: <u>AShatola</u>	DEPT: QA
APPROVED BY: <u>Fred Jones</u>	DEPT: Administration	EFFECTIVE DATE: 2021/10/26	
REVISION: 0		DCO: N/A	