

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 8784(1-3) Exp:2021/09/30

AccuLite® CLIA	87841		87842		87843	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	3.62	2.43-4.82	17.83	11.99-23.72	91.10	61.04-121.16
hsCRP in µg/ml	0.84	0.56-1.12	2.37	1.59-3.15	6.75	4.52-8.98
Troponin in ng/ml	<0.5		2.96	1.99-3.94	8.43	5.65-11.21
Myoglobin in ng/ml	29.54	19.79-39.29	31.32	20.98-41.66	59.22	39.68-78.77

DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: Research

 VERIFIED BY: *Ashatola* DEPT: QC

 APPROVED BY: *Tralagan* DEPT: Administration

EFFECTIVE DATE:2021-NOV-08

REVISION: 0

DCO: N/A