

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 8785(1-3) Exp: 2025-04-30

AccuLite® CLIA	87851		87852		87853	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	3.10	2.08-4.12	16.30	10.92-21.68	85.80	57.49-114.11
hsCRP in µg/ml	0.66	0.44-0.88	2.27	1.52-3.01	7.04	4.72-9.36
Myoglobin in ng/ml	55.03	36.87-73.19	142.12	97.23-193.01	319.97	214.38-425.55

DOCUMENT HISTORY

PREPARED BY: _____ DEPT: QC

VERIFIED BY: _____ DEPT: QA

APPROVED BY: _____ DEPT: Administration

EFFECTIVE DATE: 2022-06-24

REVISION: 0

DCO: N/A