



**VALUE ASSIGNMENT FOR BIO-RAD**
**MASTER LOT 5409(1-3) Exp date: 2024-04-30**


AccuBind® ELISA	54091		54092		54093	
	Mean	Range	Mean	Range	Mean	Range
<b>Liquichek Maternal Serum 1st Trimester</b>						
hCG in mIU/mL	1661.87	1113.45-2210.29	7666.58	5136.61-10196.55	15193.85	10179.88-20207.82
β-hCG in ng/ml	8.37	5.61-11.14	43.81	29.35-58.26	76.98	51.58-102.39
hCG-XR in mIU/mL	1739.33	1165.35-2313.30	7530.16	5045.21-10015.11	14485.66	9705.39-19265.92
PAPP-A in µg/ml	4.0	2.68-5.32	10.80	7.23-14.36	>32	

**Note 1:** The HCG data was obtained by diluting Level I , II and III at 1:1000 dilution while the HCG-XR data was obtained by diluting Level I , II and III at 1:100 dilution respectively . The dilution factor was later applied to generate the correct value for each level.

**Note 2:** PAPP-A data is being reported in µg/ml. To convert µg/ml into mIU/l multiply the values x 156.25

## DOCUMENT HISTORY

PREPARED BY:  DEPT: QC  
 APPROVED BY:  DEPT: Administration  
 REVISION: 0

VERIFIED BY:  DEPT: QA  
 EFFECTIVE DATE: 2022-07-03  
 DCO: N/A