



Neonatal 17 α -OH Progesterone (N-17-OHP) Test System
Product Code: 5525-300
Improved Performance

1.0 INTRODUCTION

Intended Use: The Quantitative Determination of 17-hydroxyprogesterone concentration in Human Whole Blood by a Microplate Enzyme Immunoassay, Colorimetric.

2.0 SUMMARY AND EXPLANATION OF THE TEST

The results are used to screen newborns for CAH (congenital adrenal hyperplasia). CAH is a genetic disorder, 90% of which is caused by 21-hydroxylase deficiency. The incidence is roughly estimated to be 1 in 15,000 newborns and can reach as high as 1 in 1480 in native Alaskans. Early diagnosis is valuable to detect CAH in newborns afflicted with the disease, not clinically recognizable, but which will lead to life threatening adrenal crisis in the neonatal period and to determine the cause of infants with ambiguous genitalia. Delayed diagnosis may also lead to further virilization in female children, acceleration of skeletal maturation and premature development of secondary sex characteristics in male children. Prompt treatment can save the life of infants and allow afflicted children to attain normal growth.


17-OHP is a steroid produced in the adrenal cortex and the gonads. It is the immediate precursor to 11-desoxycortisol (CpS), which is converted to cortisol. Because CpS is produced by 21-hydroxylation of 17-OHP, measurement of 17-OHP is an indirect indicator of 21-hydroxylase activity. CAH occurs where there is a deficiency of this enzyme. The result is a decrease in the conversion of 17-OHP to CpS which blocks the normal synthesis of cortisol. Due to the feedback mechanism, a decrease in cortisol causes an increase in ACTH secretion resulting in adrenal hyperplasia. As 17-OHP is not being converted, increased concentrations of this steroid will be found.

17-OHP concentration increases during pregnancy in the maternal and fetal blood. After birth, values decline rapidly to reach normal adult values in 2 to 7 days. Thus it is advisable not to collect samples before the 3rd day of life. Premature and sick term infants exhibit 2 to 3 fold 17-OHP values with no CAH disorder. It is suggested that a different cut off be adapted to pre-term and sick infants.

3.0 PRINCIPLE

Delayed Equilibrium Enzyme Immunoassay:


The essential reagents required for an enzyme immunoassay include antibody, enzyme-antigen conjugate and native antigen. Upon mixing the biotinylated antibody with a serum containing the antigen, a reaction results between the antigen and the antibody. The interaction is illustrated by the following equation:




Ab_(Bn) = Specific Biotinylated Antibody (Constant Quantity)
 Ag = Native Antigen (Variable Quantity)
 Ag Ab_(Bn) = Antigen-Antibody complex (Variable Quantity)

k_a = Rate Constant of Association
 k_{-a} = Rate Constant of Disassociation
 k = k_a / k_{-a} = Equilibrium Constant


After a short period, the enzyme conjugate is added (This delayed addition permits an increase in sensitivity for low concentration samples and better precision). Upon the addition of the enzyme conjugate, competition reaction results between the enzyme analog and the antigen in the sample for a limited number of antibody binding sites (not consumed in the first incubation).



Simultaneously, the immune complex is immobilized through the interaction with streptavidin coated to the well. Unbound reactants (17-OHP and 17-OHP-HRP) are removed at the end of the incubation time.



A substrate is then reacted with the enzyme bound on the wall of the microwells. The enzymatic reaction is terminated with an acid. The end product is measured at 450 nm. The 17-OHP of the unknown sample is determined using a calibration curve generated with known concentration of 17-OHP:



4.0 REAGENTS

Material Provided:

A. N-17-OHP Calibrators- Dried Blood Spots (Two rows by six dots levels - 2 x 6)

Six (6) levels of N-17-OHP calibrators in dried blood spots (adjusted to 55% hematocrit) at **approximate** concentrations of 0 (A), 6 (B), 13.5 (C), 26 (D), 53 (E) and 105 (F) ng/ml placed on WHATMAN type 903 filter paper. Store at 2-8°C. A preservative has been added.

B. N-17-OHP Controls – Dried Blood Spots (Two rows by three dots - 2 x 3)

Three (3) levels of N-17-OHP controls in dried blood spots (adjusted to 55% hematocrit) with different concentrations placed on WHATMAN type 903 filter paper in circles C1, C2, and C3. Store at 2-8°C. A preservative has been added.

Note 1: Calibrator and control values are Lot specific and were manufactured to fall within significant clinical ranges. The exact values are printed on the outside of the aluminum pouch used for storage.

Note 2: The Lot Specific calibrators, whole human blood based, were verified with the N-17-OHP blood spots supplied by CDC.

Note 3: Do not use blood spots with appearance of caking, clotting or moisture.

C. N-17-OHP Biotin Reagent – 13ml vial – Icon ▼

One (1) vial contains Anti- N-17-OHP polyclonal IgG labeled with biotin in buffer with green dye. A preservative has been added. Store at 2-8°C.

D. N-17-OHP Enzyme Reagent – 7ml/vial – Icon ☒

One (1) vial contains N-17-OHP horseradish peroxidase (HRP) conjugates in a protein-stabilizing matrix with red dye. A preservative has been added. Store at 2-8°C.

E. Streptavidin Coated Plate – 96 wells – Icon ↓

One 96-well microplate coated with streptavidin and packaged in an aluminum bag with a drying agent. Store at 2-8°C.

F. Wash Solution Concentrate – 20 ml/vial – Icon ▲

One (1) vial contains a surfactant in buffered saline. A preservative has been added. Store at 2-8°C

G. Substrate Solution – 14 ml/vial – Icon ⚡

One (1) vial contains tetramethylbenzidine (TMB) and hydrogen peroxide (H₂O₂) in buffer. Store at 2-8°C.

H. Stop Solution – 8 ml/vial – Icon ☹

One (1) vial contains strong acid (0.5M H₂SO₄). Store at 2-8°C.

I. Product Instructions

Note 1: Do not use reagents beyond the kit expiration date.

Note 2: Avoid extended exposure to heat and light. **Opened reagents are stable for sixty (60) days when stored at 2-8°C. Kit and component stability are identified on label.**

Note 3: Above reagents are for a single 96-well microplate.

4.1 Required But Not Provided:

1. Laboratory Shaker capable of 150rpm rotation.
2. Dispenser(s) for repetitive deliveries of, 0.050, 0.100 & 0.350 ml (50, 100 & 350 μ l) volumes with a precision of better than 1.5%.
3. Adjustable volume (20-200 μ l) and (200-1000 μ l) dispenser(s) for conjugate dilutions.
4. 1/8th inch hole punch.
5. Tweezers to pick up the punched spots.
6. Microplate washer or a squeeze bottle (optional).
7. Microplate Reader capable of absorbance readings at 450nm and 620nm.
8. Absorbent Paper for blotting the microplate wells.
9. Plastic wrap and microplate cover for incubation steps.
10. Vacuum aspirator (optional) for wash steps.
11. Timer.
12. External quality control

5.0 PRECAUTIONS

For In Vitro Diagnostic Use
Not for Internal or External Use in Humans or Animals

All products that contain human blood have been found to be non-reactive for Hepatitis B Surface Antigen, HIV 1&2 and HCV Antibodies by FDA required tests. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories," 2nd Edition, 1988, HHS Publication No. (CDC) 88-8395.

Safe disposal of kit components must be according to local and regulatory statutory requirement.

6.0 SPECIMEN COLLECTION AND PREPARATION

Follow the guidelines in the NCCLS publication LA47⁷ for collecting blood samples in the neonatal screening program, which can be obtained from: NCCLS, 771 E. Lancaster Ave, Villanova, PA 19085. Use WHATMAN type 903 filter paper. For samples screening for CAH, collect samples 3 to 5 days after birth. Use disposable lancets with tips less than 2.5 mm to prick the medial or lateral sides of the bottom of the heel. Allow a drop of blood to form with sufficient volume to fill a 5/8 inch diameter spot on filter paper. Gently touch the drop of blood with the filter paper. DO NOT PRESS AGAINST THE SKIN. DO NOT TOUCH SPOTTED AREA. Suspend spotted papers horizontally and allow drying at room temperature for a minimum of 3 hours. Avoid spots touching other surfaces and keep away from direct light. The samples should be transported to the laboratory within 24 hours after collection in appropriate storage container. The laboratory should store the specimens at 2-8 °C protected from moisture and direct light.

The dried blood spots are stable for at least 3 weeks at 2-8 °C protected from light and moisture. Reject samples with the following conditions:

1. Specimens not collected on WHATMAN type 903 filter paper.
2. Blood spots not completely saturated on both sides.
3. Blood spots with appearance of caking or clotting.
4. Blood spots with appearance of moisture.

7.0 QUALITY CONTROL

Each laboratory should assay controls at levels in the low, normal, and elevated range for monitoring assay performance. These controls should be treated as unknowns and values determined in every test procedure performed. Quality control charts should be maintained to follow the performance of the supplied reagents. Pertinent statistical methods should be employed to ascertain trends. Significant deviation from established performance can indicate unnoticed change in experimental conditions or degradation of kit reagents. Fresh reagents should be used to determine the reason for the variations.

8.0 REAGENT PREPARATION

1. **Wash Buffer**
Dilute contents of wash solution to 1000 ml with distilled or

deionized water in a suitable storage container. The reagent can be stored at room temperature 2-30°C for up to 60 days.

Note 1: Do not use the substrate if it looks blue.

Note 2: Do not use reagents that are contaminated or have bacteria growth.

9.0 TEST PROCEDURE

Before proceeding with the assay, bring all reagents and patient samples to room temperature (20 - 27°C).

****Test procedure should be performed by a skilled individual or trained professional****

1. Assemble the required number of microwells for each calibrator, control and patient sample to be assayed in duplicate. **Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8°C.**
 2. Punch out 1/8" blood dot out of each calibrator, control and specimens into the assigned wells. **(NOTE: Do not punch blood dots from areas that are printed or that are near the edge of the blood spot).**
 3. Add 0.100 ml (100 μ l) of N-17-OHP Biotin Reagent to all the wells.
 4. Shake the microplate for 20-30 seconds to mix. **(NOTE: Make sure that all blood dots are fully submerged in the liquid and not stuck to the walls of the microwells).**
 5. Cover with a microplate cover and rotate for 30 minutes at ambient temperature using a laboratory rotator set @ 150rpm.
 6. Remove from shaker and add 0.050 ml (50 μ l) of N-17-OHP Enzyme Reagent directly to each well. **Do not remove the reactants (DBS) in the well.**
 7. Shake the microplate gently for 20-30 seconds to mix. **(NOTE: Make sure that all blood dots are fully submerged in the liquid and not stuck to the walls of the microwells).**
 8. Cover with a microplate cover and rotate for 90 minutes at ambient temperature using a laboratory rotator set @ 150rpm.
 9. Discard the contents of the microplate by decantation or aspiration. If decanting, blot the plate dry with absorbent paper. **NOTE: Make sure all the blood dots are removed at this point. There should be no dots left in the microwells.**
 10. Add 0.350 ml (350 μ l) of wash buffer (see Reagent Preparation Section), decant (tap and blot) or aspirate. Repeat four (4) additional times for a total of five (5) washes. **An automatic or manual plate washer can be used. Follow the manufacturer's instruction for proper usage. If a squeeze bottle is employed, fill each well by depressing the container (avoiding air bubbles) to dispense the wash. Decant the wash and repeat four (4) additional times.**
 11. Add 0.100 ml (100 μ l) of substrate solution (color developer) to each well.
DO NOT SHAKE PLATE AFTER SUBSTRATE ADDITION
 12. Cover the microplate and incubate for 15 minutes at ambient temperature.
 13. Add 0.050 ml (50 μ l) of stop solution to each well and gently mix until a uniform color is obtained.
NOTE: Always add reagents in the same order to minimize reaction time differences between wells.
 14. Read the absorbance in each well at 450nm (using a reference wavelength of 620-630nm to minimize well imperfections) in a microplate reader. **The results should be read within fifteen (15) minutes of adding the stop solution.**
- Note:** Calibrators and controls are provided in the kit and should be assayed in duplicates. Each plate is limited to 38 patient samples when calibrators and controls are included. Should there be more than 38 samples needed to assay at the same time, it is advisable to assay 1 set of calibrator and control in singles per plate and take the average O.D. from each plate for calculation. Assay no more than 3 plates at a time. This means the maximum patient sample load is 129 samples.

9.1 ALTERNATIVE TEST PROCEDURE (No delay):

1. Punch out 1/8" blood dot out of each calibrator, control and specimens into the assigned wells. **(NOTE: Do not punch blood dots from areas that are printed or that are near the edge of the blood spot).**
2. Add 0.100 ml (100 μ l) of Biotin Reagent to all the wells.
3. Shake the microplate gently for 20-30 seconds to mix. **(NOTE: Make sure that all blood dots are fully submerged in the liquid and not stuck to the walls of the microwells).**
4. Add 0.050 ml (50 μ l) of Enzyme Reagent to all the wells.

- Shake the microplate gently for 20-30 seconds to mix. **(NOTE: Make sure that all blood dots are fully submerged in the liquid and not stuck to the walls of the microwells).**
- Cover with a microplate cover and rotate for 120 minutes at ambient temperature using a laboratory rotator set @ 150rpm.
- Follow steps 9-14 in the "Test Procedure" in Section 9.0.

10.0 CALCULATION OF RESULTS

A dose response curve is used to ascertain the concentration of N-17-OHP in unknown specimens.

- Record the absorbance obtained from the printout of the microplate reader as outlined in Example 1.
- Plot the absorbance for each duplicate serum reference versus the corresponding N-17-OHP concentration in ng/ml on semi-log graph paper (average the duplicates of the serum references before plotting).
- Draw the best-fit curve through the plotted points.
- To determine the concentration of N-17-OHP for an unknown, locate the average absorbance of the duplicates for each unknown on the vertical axis of the graph, find the intersecting point on the curve, and read the concentration (in ng/ml) from the horizontal axis of the graph (the duplicates of the unknown may be averaged as indicated). In the following example, the average absorbance (0.521) intersects the dose response curve at 47.0 ng/ml 17-OHP concentration.

Note: Computer data reduction software designed for ELISA assays may also be used for the data reduction. If such data reduction software is utilized, the validation of the software should be ascertained.

EXAMPLE 1 (30+90+15min Procedure)

Sample I.D.	Well Number	Abs (A)	Mean Abs (B)	Value (ng/ml)
Cal A	A1	2.330	2.324	0.0
	B1	2.318		
Cal B	C1	1.602	2.606	6.0
	D1	1.609		
Cal C	E1	1.098	1.099	13.5
	F1	1.100		
Cal D	G1	0.789	0.786	26.0
	H1	0.783		
Cal E	A2	0.483	0.466	53.0
	B2	0.452		
Cal F	C2	0.309	0.290	105.0
	D2	0.270		
Ctrl 1	E2	1.223	1.204	11.4
	F2	1.185		
Ctrl 2	G2	0.687	0.701	31.9
	H2	0.715		
Patient	A3	0.516	0.521	47.0
	B3	0.526		

*The data presented in Example 1 is for illustration only and should not be used in lieu of a dose response curve prepared with each assay.

11.0 Q.C. PARAMETERS

In order for the assay results to be considered valid the following criteria should be met:

- The absorbance (OD) of calibrator 0 ng/ml should be ≥ 1.3 .
- Four out of six quality control pools should be within the established ranges.

12.0 RISK ANALYSIS

The MSDS and Risk Analysis Form for this product are available on request from Monobind Inc.

12.1 Assay Performance

- It is important that the time of reaction in each well is held constant to achieve reproducible results.
- Pipetting of reagents should not extend beyond ten (10) minutes to avoid assay drift.
- If more than one (1) plate is used, it is recommended to repeat the dose response curve.
- The addition of substrate solution initiates a kinetic reaction, which is terminated by the addition of stop solution. Therefore,

the substrate and stop solution should be added in the same sequence to eliminate any time-deviation during reaction.

- Plate readers measure vertically. Do not touch the bottom of the wells.
- Failure to remove adhering solution adequately in the aspiration or decantation wash step(s) may result in poor replication and spurious results.
- Use components from the same lot. No intermixing of reagents from different batches.
- Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Deviation from Monobind's IFU may yield inaccurate results.
- All applicable national standards, regulations and laws, including, but not limited to, good laboratory procedures, must be strictly followed to ensure compliance and proper device usage.
- It is important to calibrate all the equipment e.g. Pipettes, Readers, Washers and/or the automated instruments used with this device, and to perform routine preventative maintenance.
- Risk Analysis - as required by CE Mark IVD Directive 98/79/EC - for this and other devices, made by Monobind, can be requested via email from Monobind@monobind.com.

12.2 Interpretation

- Measurements and interpretation of results must be performed by a skilled individual or trained professional.**
- Laboratory results alone are only one aspect for determining patient care and should not be the sole basis for therapy, particularly if the results conflict with other determinants.
- For valid test results, adequate controls and other parameters must be within the listed ranges and assay requirements.
- If test kits are altered, such as by mixing parts of different kits, which could produce false test results, or if results are incorrectly interpreted, **Monobind shall have no liability.**
- If computer controlled data reduction is used to interpret the results of the test, it is imperative that the predicted values for the calibrators fall within 10% of the assigned concentrations.
- This assay is intended solely to screen CAH in newborns. It is not to be used for confirmatory testing, monitor therapy or prenatal testing.
- The N-17-OHP blood spot screening detects only CAH caused by 21-hydroxylase deficiency, which accounts for approximately 90% of the disorder.¹ It will not detect CAH caused by deficiency of other enzymes, notably 11- β -hydroxylase deficiencies.
- Premature and infants with low birth weights tend to have higher 17-OHP values.⁶
- Samples collected prior to the second day of life tend to have higher 17-OHP values due to placental cross over.⁵
- This is a screening test, blood spots with elevated 17-OHP values should be confirmed with an extracted 17-OHP assay using serum samples.

13.0 EXPECTED RANGE OF VALUES

REPORTABLE RANGE: Analytical Range = 5 - 105 ng/ml
Samples that fall within the calibration curve should be reported as such. Samples that fall outside the calibration curve should be reported less than 5 (<5) ng/mL or greater than 105 (>105) ng/ml. Detection limit for the assay is 0.6 ng/ml.

The following is a guideline from screening programs reported in literatures.^{2,10} For assays that meet the individual lab Q.C. criteria, report results less than 22 ng/ml. Request a second specimen for values 22-35 ng/ml. Values greater than 35 ng/ml should be confirmed with an extracted 17-OHP assay using serum samples.

Since pre-term infants have 17-OHP concentrations much higher than normal full term babies, cut off level of 80 pg/disk¹¹ or 57 ng/ml serum equivalence (assuming each 3 mm disk contains 1.4 μ l of serum) has been suggested.

One hundred forty-eight normal neonatal blood spot samples free of steroid treatments obtained from a public health screening laboratory were assayed in this 17-OHP kit. Calculated mean for the assay is 12.2ng/mL with 69% of the values falling between 5-15 ng/ml.

14.0 PERFORMANCE CHARACTERISTICS

14.1 Precision

The within and between assay precisions of the N-17-OHP

Accubind® ELISA test system were determined by analyses on three different levels of dried blood controls. The number (N), mean values (X), standard deviation (σ) and coefficient of variation (C.V.) for each of these controls are presented in Table 1 and Table 2.

TABLE 1
Within Assay Precision (Values in ng/ml)

Sample	N	X	σ	C.V.
Low	18	12.2	1.02	8.4%
Mid	18	34.3	2.67	7.8%
High	18	67.2	3.58	5.3%

TABLE 2
Between Assay Precision (Values in ng/ml)

Sample	N	X	σ	C.V.
Low	10	13.2	1.19	9.0%
Mid	10	35.4	3.05	8.6%
High	10	73.2	5.23	7.1%

*As measured in duplicate.

14.2 Sensitivity

The N-17-OHP Accubind® ELISA test system has a sensitivity of 0.60 ng/ml. The sensitivity was ascertained by determining the variability of the 0 ng/ml calibrator and using the 2σ (95% certainty) statistic to calculate the minimum dose.

14.3 Accuracy

The N-17-OHP Accubind® ELISA test system was compared with a predicate 17-OHP method. Dried blood spots with concentrations from 5-80 ng/ml were used. The total number of such specimens was 54. The least square regression equation and the correlation coefficient were computed for this N-17-OHP Accubind® ELISA method in comparison with the reference method. The data obtained is displayed in Table 3.

Method	Least Square Regression Analysis	Correlation Coefficient
This Method (y)	$y = 1.03(x) + 0.95$	0.987
Reference (x)		

Only slight amounts of bias between this method and the reference method are indicated by the closeness of the values. The least square regression equation and correlation coefficient indicates good method agreement.

14.4 Specificity

The antiserum used in the assay is highly specific for the detection of 17 α -hydroxyprogesterone. The following naturally occurring steroids are spiked in a pooled human whole blood adjusted to 55% hematocrit at different concentrations. The preparation is spotted on WHATMAN type 903 filter paper, dried and assayed. The percentage indicated is the cross reactivity at 50% intercept.

Compound	Concentration (ng/ml)	% Cross Reactivity
11-desoxycortisol	78 -20,000	5.2
Progesterone	78 -20,000	4.6
17 α -hydroxypregnenolone	78 -20,000	3.7
17 α -hydroxypregnenolone sulfate	600 -10,000	1.4
Pregnenolone sulfate	78 - 20,000	<0.1
Desoxycorticosterone	20,000	<0.1
Aldosterone	20,000	<0.1
Cholesterol	20,000	<0.1
Corticosterone	20,000	<0.1
Cortisol	20,000	<0.1
Dehydroepiandrosterone	20,000	<0.1
Dihydrotestosterone	20,000	<0.1
17 α -estradiol	20,000	<0.1
17 β -estradiol	20,000	<0.1
Estrilol	20,000	<0.1
Estrone	20,000	<0.1
Testosterone	20,000	<0.1

15.0 REFERENCES

- Maria New and Lenore Levine, "Steroid 21-hydroxylase deficiency", *Pediatric and Adolescent Endocrinology*, 13: 1-46 (1984).
- Songya Pang et al, "A pilot newborn screening for congenital adrenal hyperplasia in Alaska", *Journal of Clinical Endocrinology and Metabolism*, 55:413-420, (1982).

- Hirschfeld AJ and Fleshman JK, "A usually high incidence of salt-losing congenital adrenal hyperplasia in the Alaskan Eskimo", *J Pediatrics*, 75:492 (1969).
- New ML, DuPont B, Pang S, et al, "An update on congenital adrenal hyperplasia", *Recent progress in Hormone Research*, 37:105 (1981).
- Hughes I A, et. al, "Plasma 17 α -OH Progesterone concentrations in newborn infants", *Arch. Dis. Child* 54:347-349 (1979).
- God'o B. et. al, "Plasma 17 α -OH Progesterone in full term and pre-term infants at birth and during the early neonatal period", *Hormone Research*, 15:65-71 (1981).
- National Committee for Clinical Laboratory Studies: Blood collection on filter paper for neonatal screening programs. Approved standard. NCCLS publication IA4-A Villanova, PA: NCCLS (1988).
- DPC Package Insert. I 066, July 12, 1991 .
- Westgard JO, et al, "A multi-rule chart for Quality Control", *Clin Chem* 27:493-501 (1981).
- Hofman, et al, "Direct Solid-Phase Radioimmunoassay for screening 17 α -OH Progesterone in whole blood samples from newborns", *Clin Chem*, 31:1127-1130 (1985).
- Berry J, Betts P, and Wood PJ, "The interpretation of bloodspot 17 α -hydroxyprogesterone levels in term and pre-term neonate", *Annals of Clinical Biochemistry*, 23:546-51 (1985).
- Zachmann M, et al, "Clinical and biochemical variability of congenital adrenal hyperplasia due to 11 β -hydroxylase deficiency; a study of 25 patients", *Journal of Clinical Endocrinology and Metabolism*, 56:222-229, (1983).
- "Guidelines for the Shipment of Dried Blood Spot Specimens", *Safety & Health Monograph - May 1993*, U.S. Department of Health and Human Services.

Effective date: 2017-Jun-14 Rev. 6

MP5525

DCO: 1234

Product Code: 5525-300

Item	96 (A)	192 (B)	480 (D)	960 (E)	1920 (F)
A Calibrators	1 Set	1 Set	2 Sets	4 Sets	8 Sets
B Controls	1 Set	1 Set	2 Sets	4 Sets	8 Sets
C Biotin	1 x 13 ml	2 x 13ml	1x 52 ml	2 x 52 ml	4 x 52 ml
D Enzyme	1 x 7ml	2 x 7 ml	1x 30 ml	2 x 30 ml	4 x 30 ml
E Plate	1	2	5	10	20
F Wash	1 x 20ml	1 x 20ml	1 x 60 ml	1 x 60ml	4 x 60 ml
G Substrate	1 x 14 ml	2 x 14ml	1 x 52 ml	2 x 52 ml	4 x 52 ml
H Stop	1 x 8 ml	2 x 8 ml	1 x 30 ml	2 x 30 ml	4 x 30 ml

For Orders and Inquiries, please contact

Monobind Inc.
100 North Pointe Drive
Lake Forest, CA 92630 USA


Tel: +1949.951.2665 Email: info@monobind.com
Fax: +1949.951.3539 Web: www.monobind.com

Please visit our website to learn more about our other interesting products and services.



CE REP **CEPartner4U**, Edoosdornlaan 13,
3951DB Maarn, The Netherlands
www.cepartner4u.eu

Glossary of Symbols
(EN 980/ISO 15223)



EC REP
Authorized Rep in European Country

CE
European Conformity