

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 100311(1-3)			Exp:2027/03/31			
AccuLite® CLIA	1003111		1003112		1003113	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	3.83	2.56-5.09	13.90	9.31-18.49	63.77	42.72-84.81
hsCRP in µg/ml	0.82	0.55-1.08	2.42	1.62-3.22	7.97	5.34-10.60
Myoglobin ng/ml	79.80	53.47-106.14	143.42	96.09-190.75	338.24	226.62-449.86
Troponin in ng/ml	<0.5		1.76	1.18-2.34	5.54	3.71-7.37

DOCUMENT HISTORY

 PREPARED BY: *MSW* DEPT: QC

 VERIFIED BY: *AShatola* DEPT: QA

 APPROVED BY: *Fred Lopez* DEPT: Administration

EFFECTIVE DATE: 2024/04/19

REVISION: 0

DCO: N/A