

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 100370(L,1,2,3) EXP:2027/02/28

| AccuLite® CLIA | 100370L | | 1003701 | | 1003702 | | 1003703 | |
|-----------------------------|---------|-------------|---------|--------------|---------|---------------|---------|---------------|
| | Mean | Range | Mean | Range | Mean | Range | Mean | Range |
| Liquicheck Specialty IA | | | | | | | | |
| anti-TG in IU/ml | 41.47 | 27.78-55.15 | 86.50 | 57.96-115.05 | 135.52 | 90.80-180.24 | 337.70 | 226.26-449.14 |
| anti-TPO in IU/ml | 12.20 | 8.17-16.23 | 30.74 | 20.60-40.88 | 67.44 | 45.18-89.70 | 194.65 | 130.42-258.88 |
| C-peptide in ng/ml | 0.68 | 0.46-0.90 | 0.77 | 0.51-1.02 | 3.45 | 2.31-4.59 | >9 | |
| PTH 2nd Generation in pg/ml | 18.68 | 12.51-24.84 | 23.48 | 15.73-31.23 | 181.27 | 121.45-241.08 | 608.43 | 407.65-809.22 |
| Vit-D Direct in ng/ml | 18.59 | 12.45-24.72 | 21.02 | 12.87-29.17 | 39.65 | 26.57-52.74 | 80.53 | 53.96-107.11 |

DOCUMENT HISTORY

PREPARED BY: MSW DEPT: QC
 APPROVED BY: [Signature] DEPT: Administration
 REVISION: 0

VERIFIED BY: Ashatola DEPT: QA
 EFFECTIVE DATE: 2024-04-16
 DCO: N/A