

VALUE ASSIGNMENT FOR BIO-RAD						
MASTER LOT 8789(1-3)			Exp:2027/07/31			
AccuLite® CLIA	87891		87892		87893	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck Cardiac Markers LT Plus						
CKMB in ng/ml	2.78	1.86-3.70	13.98	8.82-19.13	77.26	51.77-102.76
hsCRP in µg/ml	0.58	0.39-0.77	2.28	1.53-3.03	7.20	4.82-9.57
Myoglobin in ng/ml	53.83	36.07-71.60	155.07	103.89-206.24	372.10	249.31-494.89
Troponin in ng/ml	0.45	0.30-0.60	2.46	1.65-3.27	6.19	4.15-8.24

DOCUMENT HISTORY

PREPARED BY: *MSW* DEPT: QC
 APPROVED BY: *Fred Lopez* DEPT: Administration
 REVISION: 0

VERIFIED BY: *Ashatola* DEPT: QA
 EFFECTIVE DATE: 2024-10-11
 DCO: N/A