Monobind Inc.
The World Resource for Diagnostic Products

VALUE ASSIGNMENT FOR BIO-RAD
MASTER LOT 100194(1-3) Exp:2028/03/31

AccuBind® ELISA	100194-1		100194-2		100194-3	
	Mean	Range	Mean	Range	Mean	Range
Liquicheck AMH Control						
AMH in ng/ml	1.72	1.15-2.29	10.39	6.96-13.81	>15	

DOCUMENT HISTORY

PREPARED BY: ____ DEPT: QC

VERIFIED BY: AShatola DEPT: QA

APPROVED BY: ____ DEPT: Administration

EFFECTIVE DATE: 2026-02-17

REVISION: 0

DCO: N/A