

Advisory Notice

Date Released: 2014-July-21

Posted By: Quality Control

Effective Date: 2014-July-21

Product Affected: Releasing Agent & Stabilizing Agent

Brief Description of Advisory: Fill volumes have been increased

To whom it may concern,

A recent fill volume increase has been made to the reagents "Releasing Agent" and "Stabilization Agent" to provide technicians more product as requested by specific customers.

Reagent	Previous Volume	New Volume
Releasing Agent	12 ml	14 ml
Stabilization Agent	0.5 ml	0.7 ml

The above reagents are present as kit components for the following products:

Folate AccuBind ELISA (7525-300)

Folate AccuLite CLIA (7575-300)

Vitamin B12 AccuBind ELISA (7625-300)

Vitamin B12 AccuLite CLIA (7675-300)

The IFU for the above products has been updated accordingly.

As an official of the company, I verify this information to be true and correct.

Sincerely,

Anthony Shatola



Quality Control Director

